

डी एन ए फिंगरप्रिंटिंग एवं निदान केंद्र

(बायोटेक्नोलॉजी विज्ञान और प्रौद्योगिकी मंत्रालय एक स्वायत्त संस्थान। भारत सरकार)

CENTRE FOR DNA FINGERPRINTING AND DIAGNOSTICS

(An Autonomous Institute of Department of Biotechnology, Ministry of Science and Technology, Govt. of India)
इनर रिंग रोड Inner Ring Road, उप्पल Uppal, हैदराबाद HYDERABAD - 500039 (तेलंगाना राज्य Telangana State) भारत India

Ph. No. 040-27216026/20, Fax: 040-27216019 Email: purchase@cdfd.org.in

AWARD DETAILS FOR THE TENDER NO. CDFD/PUR/APF/2018-19/IND11814 GEL-DOCUMENTATION SYSTEM

| Mode of Tendering | Name of the Firms Responded | Make/Model Quoted | Unit Price Rs. | Status | Name of the Firm whom the Order has been placed | Basis of Selection | PO No. & Date | Order Value Rs. |
|---------------------------|--|---|-----------------------------------|--------------|---|--------------------|----------------------|--|
| Open Tender on CPP Portal | M/s. JL Technologies | OMEGA LUM G (Part No: 81-1200-00) / APLEGEN | 10,74,626.67/- (Inclusive of GST) | Qualified | M/s. Eppendorf India Limited | Lowest Bidder | 386/2018, 27.09.2018 | 15,02,550/- (Inclusive of GST) for Qty - 3Nos. |
| | M/s. Medispec (I) Ltd. | GEL DOC – It2 (Cat No. 97-0665-02) / UVP | - | Disqualified | | | | |
| | M/s. Bio-Rad Laboratories India (P) Ltd. | Gel Doc XR+ (Cat No # 1708195 / Biorad Laboratories | 5,19,734.86/- (Inclusive of GST) | Qualified | | | | |
| | M/s. Eppendorf India Ltd. | E-BOX-CX5.TS EDGE 26MX (Part No# 161126031) / Viber | 5,00,850/- (Inclusive of GST) | Qualified | | | | |
| | M/s. Tecpharma Systems Inc. | - | - | Disqualified | | | | |


Reasons for Disqualification:

1. M/s. Medispec (I) Ltd. :

- Manually controlled white lights, UV-Transilluminators, UV Timer and there is not software controlled (Automatic turn on / off) of white lights and Uv-Transilluminator for Image capture of Gels
- Manually controlled switches for white lights, UV-Transilluminators
- Manual filter changing for calorimetric & UV applications
- Epi white lights, UV-Transilluminator and also Filter changing has to operate (on / off) manually and there is not software control for filter changing & automatic turn on / off lights for Calorimetric and UV applications. The quoted model does not match our specifications and hence not considered

2. M/s. Tecpharma Sys Inc. :

- Submitted only 1 Purchase Order Copy against 2 Nos. and not meeting the Eligibility Criteria as per Tender.
- No Installation Copies enclosed
- Not meeting the Eligibility Criteria as per Tender



पूजारी-अण्डारण एव क्रय
I/C-Stores & Purchase

